

# VERIFICATION OF CALIBRATION REPORT

OCT 04 2010

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130192</u> ✓
Name <u>BRANDON LAWRENSEN</u>		ID# <u>6609</u> ✓ Date <u>9/30/10</u> ✓
A	Agency <u>JUNEAU POLICE</u>	Phone # <u>586-0600</u>
Instrument Location <u>JUNEAU POLICE DEPT</u>		
Alco S/N <u><del>130192</del> <sup>AL</sup> X172938</u> ✓ Target Value <u>.079</u> ✓ High Pressure <u>250</u>		
B	Alco Test Values <u>.081</u> ✓ 1 <sup>st</sup> Alco <u>.081</u> ✓ 2 <sup>nd</sup> Alco	
Signature <u>[Signature]</u>		
(OVER)		

BMB  
10/5/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 10<sup>th</sup> day of DEC, 2010.

[Signature] (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N 130192

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130192

SEPTEMBER 30, 2010

OPERATOR'S NAME:  
LAWRENSON/BRANDON/J  
OPERATOR'S NUMBER: 6609  
SUBJECT'S LAST NAME:

VOC  
SUBJECT'S FIRST NAME/MI :  
VOC

O.L. #: 0123456  
DEPT/AGENCY: JNU1  
CASE/REPORT: 0123456  
TEST TYPE: V  
ALCO TARGET VALUE: .079  
ALCO S/N: X172938

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.29 in  
ALCO TARGET .079 13:40  
BLANK TEST .000 13:41  
INTERNAL STANDARD VERIFIED 13:41  
ALCO TV 30.29 in .081 13:41  
BLANK TEST .000 13:42  
SUBJECT SAMPLE .000 13:42  
BLANK TEST .000 13:43  
ALCO TV 30.29 in .081 13:43  
BLANK TEST .000 13:44

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130192

SEPTEMBER 30, 2010  
TIME 13:38

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 43c

BAROMETER: 30.29 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! "\$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ &#x2113;